

## Aerobic and Anaerobic Bacteria Test Kit



**Brand:**

**Product Code:** 180-50

**Availability:** Call for availability

**Lead Time:** 35

### Description

The procedure utilizes an extinction dilution procedure to identify and count the number of organisms present. A minimum of five vials and syringes are required for each test with a twenty-eight day incubation period for anaerobic and a minimum five day period for aerobic vials.

Size: 6.5" × 6.5" × 6.5" (17 × 17 × 17 cm)  
Weight: 13 oz (368 g)